

Quality Control Certificate



Chromogranin A ELISA

REF RE53071


LOT AB398

2028-03-31
YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	0.050	-	2%
CAL B	29.0	0.195	-	8%
CAL C	95.6	0.490	-	19%
CAL D	241.7	0.835	-	33%
CAL E	964.8	2.048	-	80%
CAL F	2025.5	2.547	-	100%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target ng/mL	upper limit ng/mL
CONTROL 1	44.2	27.5	39.3	51.1
CONTROL 2	168.4	101.8	145.4	189.0



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-07-07

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl