

Quality Control Certificate



Chromogranin A ELISA

REF RE53071


LOT X401

2026-09-30
YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	0.037	-	2%
CAL B	22.0	0.106	-	4%
CAL C	87.0	0.288	-	12%
CAL D	221.9	0.780	-	32%
CAL E	782.1	1.661	-	69%
CAL F	1935.0	2.420	-	100%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target ng/mL	upper limit ng/mL
CONTROL 1	50.9	37.4	49.9	62.3
CONTROL 2	151.3	129.4	172.6	215.7




This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2024-10-30

Accepted

This document has been produced electronically and is valid without a signature.

 IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl