

Quality Control Certificate



Chromogranin A ELISA

REF RE53071


LOT Y095

2027-01-31
YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	0.057	-	3%
CAL B	26.5	0.122	-	6%
CAL C	133.8	0.334	-	16%
CAL D	221.9	0.690	-	33%
CAL E	782.1	1.252	-	60%
CAL F	2117.7	2.075	-	100%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target ng/mL	upper limit ng/mL
CONTROL 1	43.3	28.2	40.3	52.4
CONTROL 2	135.3	101.8	145.4	189.0




This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-03-25

Accepted

This document has been produced electronically and is valid without a signature.

 IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl