

Quality Control Certificate



Chromogranin A ELISA

REF RE53071


LOT Y485

2027-04-30
YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	0.018	-	1%
CAL B	25.6	0.078	-	4%
CAL C	101.0	0.311	-	15%
CAL D	260.5	0.648	-	31%
CAL E	795.8	1.248	-	59%
CAL F	2117.7	2.114	-	100%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target ng/mL	upper limit ng/mL
CONTROL 1	41.6	27.1	41.7	56.2
CONTROL 2	147.1	111.1	171.0	230.8




This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-05-26

Accepted

This document has been produced electronically and is valid without a signature.

 IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl