

Quality Control Certificate




ACTH ELISA

REF RE53081	LOT AA564	2027-08-31 YYYY-MM-DD
--------------------	------------------	---------------------------------

Final Release Results:

	Calibrator concentration pg/mL	Measured value OD	Range OD	OD/ODmax
CAL 1	0	0.069	-	2%
CAL 2	10.9	0.224	-	7%
CAL 3	33.8	0.479	-	16%
CAL 4	78.4	0.689	-	23%
CAL 5	292.7	1.766	-	59%
CAL 6	779.2	3.007	-	100%

	Concentration found pg/mL	Acceptable range		
		lower limit pg/mL	target pg/mL	upper limit pg/mL
CONTROL 1	34.8	28.0	40.1	52.1
CONTROL 2	159.0	104.6	149.4	194.2



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-12-17

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl