

Quality Control Certificate



ACTH ELISA

REF RE53081


LOT AA669

2027-11-30
YYYY-MM-DD

Final Release Results:

	Calibrator concentration pg/mL	Measured value OD	Range OD	OD/ODmax
CAL 1	0	0.007	-	0%
CAL 2	13.7	0.020	-	1%
CAL 3	51.6	0.096	-	5%
CAL 4	78.4	0.244	-	12%
CAL 5	292.7	0.780	-	37%
CAL 6	779.2	2.098	-	100%

	Concentration found pg/mL	Acceptable range		
		lower limit pg/mL	target pg/mL	upper limit pg/mL
CONTROL 1	43.0	28.0	40.1	52.1
CONTROL 2	175.6	104.6	149.4	194.2



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-08-12

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl