

Quality Control Certificate




ACTH ELISA

REF RE53081	LOT Y898	2026-10-31 YYYY-MM-DD
--------------------	-----------------	---------------------------------

Final Release Results:

	Calibrator concentration pg/mL	Measured value OD	Range OD	OD/ODmax
CAL 1	0	0.045	-	2%
CAL 2	20.3	0.139	-	6%
CAL 3	39.5	0.256	-	10%
CAL 4	79.9	0.485	-	20%
CAL 5	226.2	1.262	-	51%
CAL 6	646.4	2.478	-	100%

	Concentration found pg/mL	Acceptable range		
		lower limit pg/mL	target pg/mL	upper limit pg/mL
CONTROL 1	65.2	49.7	63.8	77.8
CONTROL 2	143.5	113.4	145.4	177.3



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-08-06

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl