Quality Control Certificate


Leptin ELISA

REF	LOT	
RE53151	45K096-26	2028-03-31 YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL 0	0	0.04	-	2%
CAL 1	2	0.17	-	8%
CAL 2	5	0.41	-	18%
CAL 3	25	1.18	-	53%
CAL 4	50	1.74	-	78%
CAL 5	100	2.24	≥ 1.20	100%

	Concentration found ng/mL	Acceptable range lower limit ng/mL	target ng/mL	upper limit ng/mL
CONTROL LOW	5.04	3.29	5.98	8.67
CONTROL HIGH	65.6	33.7	61.2	88.7



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-10-12

This document has been produced electronically and is valid without a signature.

IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl