

Quality Control Certificate



Insulin ELISA

REF RE53171


LOT 44K026

2028-02-29
YYYY-MM-DD

Final Release Results:

	Concentration μIU/mL	Measured value OD	Range OD	OD/ODmax
CAL 0	0	0.01	-	1%
CAL 1	6.25	0.07	-	4%
CAL 2	12.5	0.20	-	10%
CAL 3	25	0.51	-	27%
CAL 4	50	0.98	-	51%
CAL 5	100	1.92	≥ 1.20	100%

	Concentration found μIU/mL	Acceptable range		
		lower limit μIU/mL	target μIU/mL	upper limit μIU/mL
Biorad 40461	13.2	8.24	15.0	21.7
Biorad 40462	65.5	41.2		> 100



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-03-06

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl