

Quality Control Certificate



Renin (active) ELISA

REF RE53321


LOT 04K095

2026-09-30
YYYY-MM-DD

Final Release Results:

	Calibrator concentration pg/mL	Measured value OD	Range OD	OD/ODmax
CAL 0	0	0.08	-	3%
CAL 1	4	0.18	-	7%
CAL 2	16	0.45	-	17%
CAL 3	32	0.86	-	33%
CAL 4	64	1.62	-	63%
CAL 5	128	2.59	≥ 1.20	100%

	Concentration found pg/mL	Acceptable range		
		lower limit pg/mL	target pg/mL	upper limit pg/mL
CONTROL Low	11.7	8.00	12.4	16.7
CONTROL High	66.1	40.5	62.3	84.0



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-09-09

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl