

# Quality Control Certificate



## CYFRA 21-1 ELISA

REF RE54101


LOT 90K036

2027-03-31  
YYYY-MM-DD

### Final Release Results:

	Calibrator concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL 0	0	0.02	-	1%
CAL 1	3	0.19	-	9%
CAL 2	10	0.45	-	21%
CAL 3	25	1.09	-	50%
CAL 4	50	2.18	≥ 1.20	100%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target ng/mL	upper limit ng/mL
CONTROL LOW	4.92	4.05	5.79	7.53
CONTROL HIGH	30.9	22.0	31.4	40.8



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-03-17

**Accepted**

This document has been produced electronically and is valid without a signature.



IBL International GmbH  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)