

# Quality Control Certificate



## CA 72-4 ELISA

REF RE54111


LOT 95K036

2027-03-31  
YYYY-MM-DD

### Final Release Results:

	Concentration U/mL	Measured value OD	Range OD	OD/ODmax
CAL 0	0	0.07	-	4%
CAL 1	3	0.18	-	10%
CAL 2	20	0.46	-	25%
CAL 3	50	0.93	-	51%
CAL 4	100	1.81	≥ 1.20	100%

	Concentration found U/mL	Acceptable range		
		lower limit U/mL	target U/mL	upper limit U/mL
Control Low	3.75	2.12	3.26	4.40
Control High	62.7	35.4	54.5	73.6



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-05-29

# Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)