

# Quality Control Certificate



## FT3 ELISA

REF **RE55231**


LOT **6534A**

**2027-05-31**  
YYYY-MM-DD

### Final Release Results:

	Concentration pg/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0.0	2.576	> 1.030	100%
CAL B	1.6	2.240	> 0.895	87%
CAL C	3.3	1.663	> 0.665	65%
CAL D	6.0	0.907	> 0.365	35%
CAL E	9.5	0.595	> 0.240	23%
CAL F	19.0	0.298	> 0.120	12%

	Concentration found pg/mL	Acceptable range		
		lower limit pg/mL	target pg/mL	upper limit pg/mL
BIORAD 40461	2.64	1.74	2.49	3.23
BIORAD 40462	4.87	3.37	4.81	6.25
BIORAD 40463	8.16	4.98	7.12	9.25



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-01-13

# Accepted

This document has been produced electronically and is valid without a signature.



**IBL International GmbH**  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11

E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)

WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)