

Quality Control Certificate



FT4 ELISA

REF RE55241


LOT 6320B

2026-06-30
YYYY-MM-DD

Final Release Results:

	Concentration ng/dL	Measured value OD	Range OD	OD/ODmax
CAL A	0.00	2.492	> 0.995	100%
CAL B	0.45	2.036	> 0.815	82%
CAL C	1.30	1.355	> 0.540	54%
CAL D	2.40	0.800	> 0.320	32%
CAL E	4.60	0.326	> 0.130	13%
CAL F	7.60	0.125	> 0.050	5%

	Concentration found ng/dL	Acceptable range		
		lower limit ng/dL	target ng/dL	upper limit ng/dL
BIORAD 40461	0.84	0.47	0.76	1.05
BIORAD 40462	1.97	1.22	1.74	2.26
BIORAD 40463	3.27	2.25	3.22	4.19



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-02-06

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl