

Quality Control Certificate



Helicobacter pylori IgG ELISA

REF	RE56381	LOT	HG-511		2026-07-08
					YYYY-MM-DD

Final Release Results:

	Concentration U/mL	Result OD	Acceptable range OD
CAL A	1	0.004	0.001 - 0.150
CAL B	10	0.545	0.200 - 0.800
CAL C	25	0.998	0.450 - 1.500
CAL D	150	1.973	0.750 - 3.000

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Jörg Spangenberg

Accepted

Validation date: 2025-01-16

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibj@tecan.com
WEB: www.tecan.com/ibj