

Quality Control Certificate



Helicobacter pylori IgG ELISA

REF RE56381

LOT HG-520



2027-09-12

YYYY-MM-DD

Final Release Results:

	Concentration U/mL	Result OD	Acceptable range OD
CAL A	1	0.003	0.001 - 0.150
CAL B	10	0.482	0.200 - 0.800
CAL C	25	0.926	0.450 - 1.500
CAL D	150	1.976	0.750 - 3.000

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Jörg Spangenberg

Accepted

Validation date: 2026-04-09

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl