

Quality Control Certificate



Helicobacter pylori IgM ELISA

REF	RE56391	LOT	HM-380		2026-04-25
					YYYY-MM-DD

Final Release Results:

	Concentration U/ml	Result OD	Acceptable range OD
CAL A	1	0.021	0.001 - 0.150
CAL B	10	0.468	0.200 - 0.800
CAL C	25	1.056	0.450 - 1.500
CAL D	150	2.048	0.750 - 3.000

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by:

Dr. Jörg Spangenberg

Accepted

Validation date:

2024-10-30

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: +49 (0) 40 532891-0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl