

Quality Control Certificate



Helicobacter pylori IgM ELISA

REF RE56391

LOT HM-388



2027-07-13

YYYY-MM-DD

Final Release Results:

| | Concentration U/mL | Result OD | Acceptable range OD |
|-------|-----------------------|--------------|------------------------|
| CAL A | 1 | 0.006 | 0.001 - 0.150 |
| CAL B | 10 | 0.552 | 0.200 - 0.800 |
| CAL C | 25 | 1.044 | 0.450 - 1.500 |
| CAL D | 150 | 2.311 | 0.750 - 3.000 |

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Jörg Spangenberg

Accepted

Validation date: 2026-02-02

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl