

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Chlamydia trachomatis IgG**

Prod.No.: **RE57011**

Lot: **CHLG-212NR**

Expiry: **2026-08-31**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0070-212	2025-02	2026-08-31
CONJ	1	KG0070-212	2025-02	2026-08-31
CONTROL -	1	NG0070-212	2025-02	2026-08-31
CUT OFF	1	CG0070-212	2025-02	2026-08-31
CONTROL +	1	PG0070-212	2025-02	2026-08-31
WASH BUF 20x	1	W-1105	2024-12	2026-12-31
SOLN STOP	1	S-1036	2024-11	2026-11-30
SUB TMB	1	T-077	2022-09	2027-03-31
DIL	1	D-489	2024-10	2026-10-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.010	< 0.200 < Cut-off
CUT OFF	0.584	0.150-1.300
CONTROL +	1.172	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	21.04	11.57 - 30.51

Final Report issued: 2025-03-04

Approved by: Hoezalp
Quality Control

Approved on: 2025-03-04



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany