



**GOLD
STANDARD
DIAGNOSTICS**



Certificate of Analysis (CoA)

Product: **Chlamydia pneumoniae IgG**

Prod.No.: **RE57041**

Lot: **CHLPG-196NR**

Expiry: **2026-07-31**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0510-196	2025-01	2026-07-31
CONJ	1	KG0510-196	2025-01	2026-07-31
CONTROL -	1	NG0510-196	2025-01	2026-07-31
CUT OFF	1	CG0510-196	2025-01	2026-07-31
CONTROL +	1	PG0510-196	2025-01	2026-07-31
WASH BUF 20x	1	W-1102	2024-10	2026-10-31
SOLN STOP	1	S-1033	2024-09	2026-09-30
SUB TMB	1	T-077	2022-09	2027-03-31
DIL	1	D-489	2024-10	2026-10-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.013	< 0.200 < Cut-off
CUT OFF	0.735	0.150-1.300
CONTROL +	1.306	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	19.90	10.95 - 28.86

Final Report issued: 2025-01-31

Approved by: _____
Quality Control

Approved on: 2025-01-31



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany