

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Chlamydia pneumoniae IgG**

Prod.No.: **RE57041**

Lot: **CHLPG-204-1**

Expiry: **2027-10-31**

IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0510-204	2026-04	2027-10-31
CONJ	1	KG0510-204	2026-04	2027-10-31
CONTROL -	1	NG0510-204	2026-04	2027-10-31
CUT OFF	1	CG0510-204-1	2026-04	2027-10-31
CONTROL +	1	PG0510-204	2026-04	2027-10-31
WASH BUF 20x	1	W-1141	2026-04	2028-04-30
SOLN STOP	1	S-1073	2026-04	2028-04-30
SUB TMB	1	T-080	2026-02	2029-02-23
DIL	1	D-501	2026-03	2028-03-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.025	< 0.200 < Cut-off
CUT OFF	0.594	0.150-1.300
CONTROL +	1.097	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	20.25	11.14 - 29.36

The target value and target range of **CONTROL +** may be used optionally.

The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2026-06-10

Approved by: Hoozalp
Quality Control

Approved on: 2026-06-10



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany