Certificate of Analysis (CoA)

Product:	**Chlamydia pneumoniae IgM**
Prod.No.:	**RE57051**
Lot:	**CHLPM-198N**
Expiry:	**2026-10-31**
IFU:	**Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0510-198	2025-04	2026-10-31
CONJ	1	KM0510-198	2025-04	2026-10-31
CONTROL -	1	NM0510-198	2025-04	2026-10-31
CUT OFF	1	CM0510-198	2025-04	2026-10-31
CONTROL +	1	PM0510-198	2025-04	2026-10-31
WASH BUF 20x	1	W-1108	2025-02	2027-02-28
SOLN STOP	1	S-1039	2025-01	2027-01-31
SUB TMB	1	T-077	2022-09	2027-03-31
DIL M	1	DM-343	2024-12	2026-12-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.010	< 0.200 < Cut-off
CUT OFF	0.460	0.150-1.300
CONTROL +	1.039	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	23.05	12.68 - 33.42

The target value and target range of CONTROL + may be used optionally.

The positive control is valid if the absorbance value of CONTROL + > CUT OFF.

Final Report issued: 2025-04-16

Approved by: _____

Quality Control

Approved on: 2025-04-16

Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany

IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany