

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Chlamydia pneumoniae IgM**

Prod.No.: **RE57051**

Lot: **CHLPM-206**

Expiry: **2027-12-31**

IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0510-206	2026-06	2027-12-31
CONJ	1	KM0510-206	2026-06	2027-12-31
CONTROL -	1	NM0510-206	2026-06	2027-12-31
CUT OFF	1	CM0510-206	2026-06	2027-12-31
CONTROL +	1	PM0510-206	2026-06	2027-12-31
WASH BUF 20x	1	W-1143	2026-05	2028-05-31
SOLN STOP	1	S-1075	2026-04	2028-04-30
SUB TMB	1	T-080	2026-02	2029-02-23
DIL M	1	DM-354	2026-04	2028-04-30

Component	OD Value	OD Validation Criteria
CONTROL -	0.061	< 0.200 < Cut-off
CUT OFF	0.751	0.150-1.300
CONTROL +	1.666	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	21.04	11.57 - 30.50

The target value and target range of **CONTROL +** may be used optionally.

The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2026-06-23

Approved by: 
Quality Control

Approved on: 2026-06-23



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany