

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Cytomegalovirus (CMV) IgG**
 Prod.No.: **RE57061**
 Lot: **CMVG-138**
 Expiry: **2026-11-30**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0110-138	2025-05	2026-11-30
CONJ	1	KG0110-138	2025-05	2026-11-30
CONTROL -	1	NG0110-138	2025-05	2026-11-30
CUT OFF	1	CG0110-138	2025-05	2026-11-30
CONTROL +	1	PG0110-138	2025-05	2026-11-30
WASH BUF 20x	1	W-1110	2025-03	2027-03-31
SOLN STOP	1	S-1041	2025-02	2027-02-28
SUB TMB	1	T-077	2022-09	2027-03-31
DIL	1	D-491	2025-02	2027-02-28

Component	OD Value	OD Validation Criteria
CONTROL -	0.004	< 0.200 < Cut-off
CUT OFF	0.459	0.150-1.300
CONTROL +	1.144	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	22.08	12.14 - 32.01

The target value and target range of **CONTROL +** may be used optionally.
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2025-06-02

Approved by: 
 Quality Control

Approved on: 2025-06-03



Gold Standard Diagnostics Frankfurt GmbH
 Waldstrasse 23 A6
 63128 Dietzenbach, Germany



IBL International GmbH
 Flughafenstrasse 52a, 22335 Hamburg
 Germany