



**GOLD
STANDARD
DIAGNOSTICS**



Certificate of Analysis (CoA)

Product: **Cytomegalovirus (CMV) IgG**
 Prod.No.: **RE57061**
 Lot: **CMVG-140**
 Expiry: **2027-05-31**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0110-140	2025-11	2027-05-31
CONJ	1	KG0110-140	2025-11	2027-05-31
CONTROL -	1	NG0110-140	2025-11	2027-05-31
CUT OFF	1	CG0110-140	2025-11	2027-05-31
CONTROL +	1	PG0110-140	2025-11	2027-05-31
WASH BUF 20x	1	W-1128	2025-11	2027-11-30
SOLN STOP	1	S-1059	2025-10	2027-10-31
SUB TMB	1	T-078	2025-04	2028-04-28
DIL	1	D-496	2025-09	2027-09-30

Component	OD Value	OD Validation Criteria
CONTROL -	0.026	< 0.200 < Cut-off
CUT OFF	0.379	0.150-1.300
CONTROL +	0.768	> Cut-off


Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	21.05	11.58 - 30.52


The target value and target range of **CONTROL +** may be used optionally.
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2025-12-03

Approved by: 
 Quality Control

Approved on: 2025-12-03

 Gold Standard Diagnostics Frankfurt GmbH
 Waldstrasse 23 A6
 63128 Dietzenbach, Germany

 IBL International GmbH
 Flughafenstrasse 52a, 22335 Hamburg
 Germany