


Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Rubella Virus IgG**

Prod.No.: **RE57081**

Lot: **RUBG-076N**

Expiry: **2026-07-31**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0400-076	2025-01	2026-07-31
CONJ	1	KG0400-076	2025-01	2026-07-31
CAL A	1	SAG0400-076	2025-01	2026-07-31
CAL B	1	SBG0400-076	2025-01	2026-07-31
CAL C	1	SCG0400-076	2025-01	2026-07-31
CAL D	1	SDG0400-076	2025-01	2026-07-31
WASH BUF 20x	1	W-1103	2024-10	2026-10-31
SOLN STOP	1	S-1034	2024-10	2026-10-31
SUB TMB	1	T-077	2022-09	2027-03-31
DIL	1	D-489	2024-10	2026-10-31

Component	Concentration	OD Value	OD Validation Criteria
CAL A	0 IU/mL	0.009	< 0.200
CAL B	10 IU/mL	0.999	> 0.200 & > Standard A
CAL C	50 IU/mL	1.947	> 0.700 & > Standard B
CAL D	100 IU/mL	2.434	> 1.100 & > Standard C

Final Report issued: 2025-01-30

Approved by: *S. Herrmann*
Quality Control



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany

Approved on: 2025-01-31



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany