

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Rubella Virus IgG**
 Prod.No.: **RE57081**
 Lot: **RUBG-078**
 Expiry: **2027-01-31**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0400-078	2026-07	2027-01-31
CONJ	1	KG0400-078	2026-07	2027-01-31
CAL A	1	SAG0400-078	2026-07	2027-01-31
CAL B	1	SBG0400-078	2026-07	2027-01-31
CAL C	1	SCG0400-078	2026-07	2027-01-31
CAL D	1	SDG0400-078	2026-07	2027-01-31
WASH BUF 20x	1	W-1114	2025-05	2027-05-31
SOLN STOP	1	S-1045	2025-04	2027-04-30
SUB TMB	1	T-078	2025-04	2028-04-28
DIL	1	D-492	2025-03	2027-03-31

Component	Concentration	OD Value	OD Validation Criteria
CAL A	0 IU/mL	0.007	< 0.200
CAL B	10 IU/mL	0.362	> 0.200 & > Standard A
CAL C	50 IU/mL	0.934	> 0.700 & > Standard B
CAL D	100 IU/mL	1.595	> 1.100 & > Standard C

Final Report issued: 2025-07-16

Approved by: 
 Quality Control

Approved on: 2025-07-16