

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Rubella Virus IgM μ -capture**

Prod.No.: **RE57091**

Lot: **RUBM-168N**

Expiry: **2026-07-31**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0400-168	2025-01	2026-07-31
CONJS	1	KM0400-168	2025-01	2026-07-31
CONTROL -	1	NM0400-168	2025-01	2026-07-31
CUT OFF	1	CM0400-168	2025-01	2026-07-31
CONTROL +	1	PM0400-168	2025-01	2026-07-31
WASH BUF 20x	1	W-1102	2024-10	2026-10-31
SOLN STOP	1	S-1033	2024-09	2026-09-30
SUB TMB	1	T-077	2022-09	2027-03-31
DIL	1	D-488	2024-08	2026-08-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.078	< Cut-off
CUT OFF	0.422	0.150-1.300
CONTROL +	1.021	> Cut-off

Final Report issued: 2025-01-24

Approved by: *HOe*
Quality Control

Approved on: 2025-01-24



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany