

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Toxoplasma gondii IgM μ -capture**

Prod.No.: **RE57111**

Lot: **TOXM-245**

Expiry: **2027-02-28**

IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0460-245	2025-08	2027-02-28
CONJS	1	KM0460-245	2025-08	2027-02-28
CONTROL -	1	NM0460-245	2025-08	2027-02-28
CUT OFF	1	CM0460-245	2025-08	2027-02-28
CONTROL +	1	PM0460-245	2025-08	2027-02-28
WASH BUF 20x	1	W-1117	2025-07	2027-07-31
SOLN STOP	1	S-1048	2025-05	2027-05-31
SUB TMB	1	T-078	2025-04	2028-04-28
DIL	1	D-493	2025-05	2027-05-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.031	< Cut-off
CUT OFF	0.467	0.150-1.300
CONTROL +	1.086	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	25.44	13.99 - 36.89

The target value and target range of **CONTROL +** may be used optionally.

The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2025-09-05

Approved by: 
Quality Control

Approved on: 2025-09-05



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany