



**GOLD
STANDARD
DIAGNOSTICS**



Certificate of Analysis (CoA)

Product: **Parvovirus B19 IgG**
 Prod.No.: **RE57281**
 Lot: **PARG-200N**
 Expiry: **2026-07-31**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0370-200	2025-01	2026-07-31
CONJ	1	KG0370-200	2025-01	2026-07-31
CONTROL -	1	NG0370-200	2025-01	2026-07-31
CUT OFF	1	CG0370-200	2025-01	2026-07-31
CONTROL +	1	PG0370-200	2025-01	2026-07-31
WASH BUF 20x	1	W-1104	2024-12	2026-12-31
SOLN STOP	1	S-1035	2024-10	2026-10-31
SUB TMB	1	T-077	2022-09	2027-03-31
DIL	1	D-489	2024-10	2026-10-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.012	< 0.200 < Cut-off
CUT OFF	0.750	0.150-1.300
CONTROL +	1.696	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	21.47	11.81 - 31.12

The target value and target range of **CONTROL +** may be used optionally.
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2025-02-20

Approved by: 
 Quality Control

Approved on: 2025-02-20



Gold Standard Diagnostics Frankfurt GmbH
 Waldstrasse 23 A6
 63128 Dietzenbach, Germany



IBL International GmbH
 Flughafenstrasse 52a, 22335 Hamburg
 Germany