

# Quality Control Certificate



## TBE virus (FSME) IgG ELISA

<b>REF</b> RE57401	<b>LOT</b> IFSG135	<b>2026-05-31</b> YYYY-MM-DD
--------------------	--------------------	---------------------------------

### Final Release Results:

	Concentration VIEU/mL	Measured value OD	manual		e.g. automated/DSX Measured value OD
			Acceptable range*		
			lower limit OD	upper limit OD	
CAL 1	50	0.182		0.500	0.133
CAL 2	110	0.326			0.242
CAL 3	250	0.573			0.473
CAL 4	550	1.401			1.180
CAL 5	900	2.056	1.000		2.068

\*During automated procedure the optical densities of calibrators and controls may rise due to temperature and specific device conditions.  
In case of none conformance with the OD ranges stated on the QC Certificate please contact IBL-International or use the validation criteria only!

	Concentration found VIEU/mL	Validation criteria		
		lower limit VIEU/mL	target VIEU/mL	upper limit VIEU/mL
CONTROL LL	156	109	156	203
CONTROL HL	370	281	402	522

### Kit Composition:

REF		LOT		REF		LOT	
KIFSG11	MTP	L330	2026-07	KIZZZ71	TMB STOP	L205	2027-03
KIFSG41	ENZCONJ	L332	2026-05				
KIFSG51	CAL 1-5	M788A-2	2026-09				
KIFSG61	CONTROL LL	L352	2027-05				
KIFSG71	CONTROL HL	L353	2027-05				
KIZZZ61	DILBUF	L331	2029-04				
KIZZZ01	WASHBUF	L310	2027-10				
KEZZ991	TMB SUBS	M840-4	2027-02				

Valid edition of the instructions for use: 2016-06

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: S. Geweke

Validation date: 2024-06-18

**Accepted**

This document has been produced electronically and is valid without a signature.



**IBL International GmbH**  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)