

Quality Control Certificate



TBE virus (FSME) IgG ELISA

REF RE57401	LOT IFSG136	2026-10-31 YYYY-MM-DD
--------------------	--------------------	---------------------------------

Final Release Results:

	Concentration VIEU/mL	Measured value OD	manual		e.g. automated/DSX Measured value OD
			Acceptable range*		
			lower limit OD	upper limit OD	
CAL 1	65	0.178		0.500	0.117
CAL 2	140	0.348			0.217
CAL 3	260	0.690			0.432
CAL 4	540	1.498			1.013
CAL 5	1100	2.986	1.000		1.987

*During automated procedure the optical densities of calibrators and controls may rise due to temperature and specific device conditions.
In case of none conformance with the OD ranges stated on the QC Certificate please contact IBL-International or use the validation criteria only!

	Concentration found VIEU/mL	Validation criteria		
		lower limit VIEU/mL	target VIEU/mL	upper limit VIEU/mL
CONTROL LL	176	140	200	260
CONTROL HL	454	347	496	645

Kit Composition:

REF		LOT		REF		LOT	
KIFSG11	MTP	L649	2027-01	KIZZZ71	TMB STOP	L205	2027-03
KIFSG41	ENZCONJ	L674	2026-10				
KIFSG51	CAL 1-5	L657	2027-10				
KIFSG61	CONTROL LL	L658	2027-10				
KIFSG71	CONTROL HL	L659	2027-10				
KIZZZ61	DILBUF	L605	2029-09				
KIZZZ01	WASHBUF	L606	2028-03				
KEZZ991	TMB SUBS	L541-3	2027-10				

Valid edition of the instructions for use: 2016-06

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: M.Dries

Validation date: 2025-01-14

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl