

Quality Control Certificate



TBE virus (FSME) IgG ELISA

REF RE57401	LOT IFSG138	2028-04-30 YYYY-MM-DD
--------------------	--------------------	---------------------------------

Final Release Results:

	Concentration VIEU/mL	Measured value OD	manual		e.g. automated/DSX Measured value OD
			Acceptable range*		
			lower limit OD	upper limit OD	
CAL 1	60	0.157	1.000	0.500	0.155
CAL 2	120	0.313			0.304
CAL 3	250	0.581			0.561
CAL 4	600	1.464			1.514
CAL 5	1100	2.361			2.513

*During automated procedure the optical densities of calibrators and controls may rise due to temperature and specific device conditions.
In case of none conformance with the OD ranges stated on the QC Certificate please contact IBL-International or use the validation criteria only!

	Concentration found VIEU/mL	Validation criteria		
		lower limit VIEU/mL	target VIEU/mL	upper limit VIEU/mL
CONTROL LL	151	111	158	206
CONTROL HL	404	292	417	542

Kit Composition:

REF		LOT		REF		LOT	
KIFSG11	MTP	K704	2028-06	KIZZZ71	TMB STOP	K646	2029-02
KIFSG41	ENZCONJ	K735	2028-04				
KIFSG51	CAL 1-5	K666	2029-03				
KIFSG61	CONTROL LL	K667	2029-03				
KIFSG71	CONTROL HL	K668	2029-03				
KIZZZ61	DILBUF	K665	2031-03				
KIZZZ01	WASHBUF	K645	2029-08				
KEZZ991	TMB SUBS	K610-2	2029-08				

Valid edition of the instructions for use: 2016-06

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T. Borchert

Validation date: 2026-05-26

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl