

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Herpes simplex Virus 2 (HSV 2) IgG**

Prod.No.: **RE57921**

Lot: **HSV2G-146**

Expiry: **2027-09-30**

IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0540-146	2026-03	2027-09-30
CONJ	1	KG0540-146	2026-03	2027-09-30
CONTROL -	1	NG0540-146	2026-03	2027-09-30
CUT OFF	1	CG0540-146	2026-03	2027-09-30
CONTROL +	1	PG0540-146	2026-03	2027-09-30
WASH BUF 20x	1	W-1138	2026-03	2028-03-31
SOLN STOP	1	S-1070	2026-02	2028-02-29
SUB TMB	1	T-079	2025-12	2028-12-10
DIL	1	D-500	2026-02	2028-02-29


Component	OD Value	OD Validation Criteria
CONTROL -	0.015	< 0.200 < Cut-off
CUT OFF	0.463	0.150-1.300
CONTROL +	1.121	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	23.73	13.05 - 34.40

The target value and target range of **CONTROL +** may be used optionally.

The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2026-04-13

Approved by: 
Quality Control

Approved on: 2026-04-13



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany