

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Herpes simplex Virus 2 (HSV 2) IgM**

Prod.No.: **RE57931**

Lot: **HSV2M-140N**

Expiry: **2026-09-30**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0540-140	2025-03	2026-09-30
CONJ	1	KM0540-140	2025-03	2026-09-30
CONTROL -	1	NM0540-140	2025-03	2026-09-30
CUT OFF	1	CM0540-140	2025-03	2026-09-30
CONTROL +	1	PM0540-140	2025-03	2026-09-30
WASH BUF 20x	1	W-1107	2025-01	2027-01-31
SOLN STOP	1	S-1038	2024-12	2026-12-31
SUB TMB	1	T-077	2022-09	2027-03-31
DIL M	1	DM-343	2024-12	2026-12-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.013	< 0.200 < Cut-off
CUT OFF	0.633	0.150-1.300
CONTROL +	1.452	> Cut-off

Final Report issued: 2025-03-28

Approved by: Hoezalp
Quality Control

Approved on: 2025-03-28



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany