

# Certificate of Analysis (CoA)



**GOLD  
STANDARD  
DIAGNOSTICS**



Product: **Herpes simplex Virus 1+2 (HSV) IgG**

Prod.No.: **RE57941**

Lot: **HSVG-113N**

Expiry: **2026-06-30**

Component	Number	Lot	Prod. Date	Expiry Date
<b>MTP</b>	1	MG0250-113	2024-12	2026-06-30
<b>CONJ</b>	1	KG0250-113	2024-12	2026-06-30
<b>CONTROL -</b>	1	NG0250-113	2024-12	2026-06-30
<b>CUT OFF</b>	1	CG0250-113	2024-12	2026-06-30
<b>CONTROL +</b>	1	PG0250-113	2024-12	2026-06-30
<b>WASH BUF 20x</b>	1	W-1101	2024-10	2026-10-31
<b>SOLN STOP</b>	1	S-1032	2024-09	2026-09-30
<b>SUB TMB</b>	1	T-077	2022-09	2027-03-31
<b>DIL</b>	1	D-488	2024-08	2026-08-31

Component	OD Value	OD Validation Criteria
<b>CONTROL -</b>	0.012	< 0.200 < Cut-off
<b>CUT OFF</b>	0.720	0.150-1.300
<b>CONTROL +</b>	1.381	> Cut-off

Final Report issued: 2025-01-14

Approved by:   
Quality Control

Approved on: 2025-01-14



Gold Standard Diagnostics Frankfurt GmbH  
Waldstrasse 23 A6  
63128 Dietzenbach, Germany



IBL International GmbH  
Flughafenstrasse 52a, 22335 Hamburg  
Germany