

**GOLD
STANDARD
DIAGNOSTICS**



Certificate of Analysis (CoA)

Product: **Herpes simplex Virus 1+2 (HSV) IgM**
 Prod.No.: **RE57951**
 Lot: **HSVM-119**
 Expiry: **2027-06-30**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0250-119	2025-12	2027-06-30
CONJ	1	KM0250-119	2025-12	2027-06-30
CONTROL -	1	NM0250-119	2025-12	2027-06-30
CUT OFF	1	CM0250-119	2025-12	2027-06-30
CONTROL +	1	PM0250-119	2025-12	2027-06-30
WASH BUF 20x	1	W-1129	2025-11	2027-11-30
SOLN STOP	1	S-1064	2025-11	2027-11-30
SUB TMB	1	T-078	2025-04	2028-04-28
DIL M	1	DM-350	2025-12	2027-12-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.005	< 0.200 < Cut-off
CUT OFF	0.520	0.150-1.300
CONTROL +	1.075	> Cut-off


Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	21.03	11.56 - 30.49


The target value and target range of **CONTROL +** may be used optionally.
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2026-01-06

Approved by: 
 Quality Control

Approved on: 2026-01-06

 Gold Standard Diagnostics Frankfurt GmbH
 Waldstrasse 23 A6
 63128 Dietzenbach, Germany

 IBL International GmbH
 Flughafenstrasse 52a, 22335 Hamburg
 Germany