



**GOLD  
STANDARD  
DIAGNOSTICS**



# Certificate of Analysis (CoA)

Product: **Legionella pneumophila IgM**  
 Prod.No.: **RE58511**  
 Lot: **LEGM-076**  
 Expiry: **2027-06-30**  
 IFU: **Rev. 1**

Component	Number	Lot	Prod. Date	Expiry Date
<b>MTP</b>	1	MM0650-076	2025-12	2027-06-30
<b>CONJ</b>	1	KM0650-076	2025-12	2027-06-30
<b>CONTROL -</b>	1	NM0650-076	2025-12	2027-06-30
<b>CUT OFF</b>	1	CM0650-076	2025-12	2027-06-30
<b>CONTROL +</b>	1	PM0650-076	2025-12	2027-06-30
<b>WASH BUF 20x</b>	1	W-1129	2025-11	2027-11-30
<b>SOLN STOP</b>	1	S-1064	2025-11	2027-11-30
<b>SUB TMB</b>	1	T-078	2025-04	2028-04-28
<b>DIL M</b>	1	DM-350	2025-12	2027-12-31

Component	OD Value	OD Validation Criteria
<b>CONTROL -</b>	0.019	< 0.200 < Cut-off
<b>CUT OFF</b>	0.678	0.150-1.300
<b>CONTROL +</b>	1.275	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
<b>CONTROL +</b>	20.91	11.50 - 30.32

The target value and target range of **CONTROL +** may be used optionally.  
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2026-01-28

Approved by:   
 Quality Control

Approved on: 2026-01-28



Gold Standard Diagnostics Frankfurt GmbH  
 Waldstrasse 23 A6  
 63128 Dietzenbach, Germany



IBL International GmbH  
 Flughafenstrasse 52a, 22335 Hamburg  
 Germany