

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Leishmania infantum IgG**

Prod.No.: **RE58641**

Lot: **LEIG-102**

Expiry: **2027-06-30**

IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0310-102	2025-12	2027-06-30
CONJ	1	KG0310-102	2025-12	2027-06-30
CONTROL -	1	NG0310-102	2025-12	2027-06-30
CUT OFF	1	CG0310-102	2025-12	2027-06-30
CONTROL +	1	PG0310-102	2025-12	2027-06-30
WASH BUF 20x	1	W-1127	2025-11	2027-11-30
SOLN STOP	1	S-1062	2025-11	2027-11-30
SUB TMB	1	T-078	2025-04	2028-04-28
DIL	1	D-496	2025-09	2027-09-30

Component	OD Value	OD Validation Criteria
CONTROL -	0.017	< 0.200 < Cut-off
CUT OFF	0.567	0.150-1.300
CONTROL +	1.106	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	22.98	12.64 - 33.31

The target value and target range of **CONTROL +** may be used optionally.

The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2025-12-15

Approved by: Hoezalp
Quality Control

Approved on: 2025-12-15



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany