



**GOLD
STANDARD
DIAGNOSTICS**



Certificate of Analysis (CoA)

Product: **Trichinella spiralis IgG**
 Prod.No.: **RE58661**
 Lot: **TRIG-085N**
 Expiry: **2026-09-30**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0480-085	2025-03	2026-09-30
CONJ	1	KG0480-085	2025-03	2026-09-30
CONTROL -	1	NG0480-085	2025-03	2026-09-30
CUT OFF	1	CG0480-085	2025-03	2026-09-30
CONTROL +	1	PG0480-085	2025-03	2026-09-30
WASH BUF 20x	1	W-1107	2025-01	2027-01-31
SOLN STOP	1	S-1038	2024-12	2026-12-31
SUB TMB	1	T-077	2022-09	2027-03-31
DIL	1	D-490	2024-12	2026-12-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.012	< 0.200 < Cut-off
CUT OFF	0.568	0.150-1.300
CONTROL +	1.147	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	24.37	13.40 - 35.34

The target value and target range of **CONTROL +** may be used optionally.
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2025-03-25

Approved by: ACr
 Quality Control

Approved on: 2025-03-25



Gold Standard Diagnostics Frankfurt GmbH
 Waldstrasse 23 A6
 63128 Dietzenbach, Germany



IBL International GmbH
 Flughafenstrasse 52a, 22335 Hamburg
 Germany