

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Dengue Virus IgG**

Prod.No.: **RE58671**

Lot: **DENG-179**

Expiry: **2026-11-30**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0120-179	2025-05	2026-11-30
CONJ	1	KG0120-179	2025-05	2026-11-30
CONTROL -	1	NG0120-179	2025-05	2026-11-30
CUT OFF	1	CG0120-179	2025-05	2026-11-30
CONTROL +	1	PG0120-179	2025-05	2026-11-30
WASH BUF 20x	1	W-1112	2025-03	2027-03-31
SOLN STOP	1	S-1043	2025-03	2027-03-31
SUB TMB	1	T-077	2022-09	2027-03-31
DIL	1	D-491	2025-02	2027-02-28

Component	OD Value	OD Validation Criteria
CONTROL -	0.020	< 0.200 < Cut-off
CUT OFF	0.545	0.150-1.300
CONTROL +	1.099	> Cut-off

Final Report issued: 2025-06-06

Approved by: 
Quality Control

Approved on: 2025-06-06



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany