

# Certificate of Analysis (CoA)



**GOLD  
STANDARD  
DIAGNOSTICS**



Product: **Dengue Virus IgM**  
 Prod.No.: **RE58681**  
 Lot: **DENM-181**  
 Expiry: **2027-08-31**  
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
<b>MTP</b>	1	MM0120-181	2026-02	2027-08-31
<b>CONJ</b>	1	KM0120-181	2026-02	2027-08-31
<b>CONTROL -</b>	1	NM0120-181	2026-02	2027-08-31
<b>CUT OFF</b>	1	CM0120-181	2026-02	2027-08-31
<b>CONTROL +</b>	1	PM0120-181	2026-02	2027-08-31
<b>WASH BUF 20x</b>	1	W-1135	2026-02	2028-02-29
<b>SOLN STOP</b>	1	S-1067	2026-01	2028-01-31
<b>SUB TMB</b>	1	T-079	2025-12	2028-12-10
<b>DIL M</b>	1	DM-351	2026-01	2028-01-31

Component	OD Value	OD Validation Criteria
<b>CONTROL -</b>	0.002	< 0.200 < Cut-off
<b>CUT OFF</b>	0.549	0.150-1.300
<b>CONTROL +</b>	1.072	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
<b>CONTROL +</b>	25.32	13.92 - 36.71

The target value and target range of **CONTROL +** may be used optionally.  
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2026-03-03

Approved by:   
 Quality Control

Approved on: 2026-03-03



Gold Standard Diagnostics Frankfurt GmbH  
 Waldstrasse 23 A6  
 63128 Dietzenbach, Germany



IBL International GmbH  
 Flughafenstrasse 52a, 22335 Hamburg  
 Germany