

Quality Control Certificate



Chagas (Trypanosoma cruzi) IgG ELISA

REF RE58691

LOT CHAG-071

 2027-02-28
YYYY-MM-DD

Final Release Results:

| | Result OD | Acceptable range OD |
|----------------------|--------------|------------------------|
| IgG Negative Control | 0.013 | < 0.200 < Cut-Off |
| IgG Positive Control | 1.440 | > Cut-Off |
| Cut off Control | 0.530 | 0.150 - 1.300 |

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-09-16

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.:
E-MAIL:
WEB:

+ 49 (0) 40 532891 -0 Fax: -11
ibl@tecan.com
www.tecan.com/ibl