

Quality Control Certificate



Chagas (Trypanosoma cruzi) IgG ELISA

REF RE58691

LOT CHAG-074

 2027-10-31
YYYY-MM-DD

Final Release Results:

	Result OD	Acceptable range OD
IgG Negative Control	0.014	< 0.200 < Cut-Off
IgG Positive Control	1.494	> Cut-Off
Cut off Control	0.585	0.150 - 1.300

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-05-18

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.:
E-MAIL:
WEB:

+ 49 (0) 40 532891 -0 Fax: -11
ibl@tecan.com
www.tecan.com/ibl