

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Ascaris lumbricoides IgG**

Prod.No.: **RE58701**

Lot: **ASCG-077N**

Expiry: **2026-07-31**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0020-077	2025-01	2026-07-31
CONJ	1	KG0020-077	2025-01	2026-07-31
CONTROL -	1	NG0020-077	2025-01	2026-07-31
CUT OFF	1	CG0020-077	2025-01	2026-07-31
CONTROL +	1	PG0020-077	2025-01	2026-07-31
WASH BUF 20x	1	W-1102	2024-10	2026-10-31
SOLN STOP	1	S-1033	2024-09	2026-09-30
SUB TMB	1	T-077	2022-09	2027-03-31
DIL	1	D-488	2024-08	2026-08-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.012	< 0.200 < Cut-off
CUT OFF	0.607	0.150-1.300
CONTROL +	1.263	> Cut-off

Final Report issued: 2025-01-28

Approved by: 
Quality Control

Approved on: 2025-01-28



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany