

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Toxocara canis IgG**

Prod.No.: **RE58721**

Lot: **TOCG-096N**

Expiry: **2026-05-31**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0450-096	2024-11	2026-05-31
CONJ	1	KG0450-096	2024-11	2026-05-31
CONTROL -	1	NG0450-096	2024-11	2026-05-31
CUT OFF	1	CG0450-096	2024-11	2026-05-31
CONTROL +	1	PG0450-096	2024-11	2026-05-31
WASH BUF 20x	1	W-1099	2024-09	2026-09-30
SOLN STOP	1	S-1030	2024-09	2026-09-30
SUB TMB	1	T-076	2022-05	2026-11-30
DIL	1	D-487	2024-07	2026-07-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.023	< 0.200 < Cut-off
CUT OFF	0.549	0.150-1.300
CONTROL +	1.168	> Cut-off

Final Report issued: 2024-11-28

Approved by: ACr
Quality Control

Approved on: 2024-11-28



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany