

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Taenia solium IgG**

Prod.No.: **RE58731**

Lot: **TAEG-087**

Expiry: **2028-01-31**

IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0420-087	2026-07	2028-01-31
CONJ	1	KG0420-087	2026-07	2028-01-31
CONTROL -	1	NG0420-087	2026-07	2028-01-31
CUT OFF	1	CG0420-087	2026-07	2028-01-31
CONTROL +	1	PG0420-087	2026-07	2028-01-31
WASH BUF 20x	1	W-1147	2026-06	2028-06-30
SOLN STOP	1	S-1080	2026-07	2028-07-31
SUB TMB	1	T-080	2026-02	2029-02-23
DIL	1	D-503	2026-06	2028-06-30


Component	OD Value	OD Validation Criteria
CONTROL -	0.007	< 0.200 < Cut-off
CUT OFF	0.517	0.150-1.300
CONTROL +	0.873	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	17.72	> 10 - 25.69

The target value and target range of **CONTROL +** may be used optionally.

The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2026-08-25

Approved by: 
Quality Control

Approved on: 2026-08-25



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany