Certificate of Analysis (CoA)

Product: **Epstein-Barr Virus (EBNA) IgG**
Prod.No.: **RE58741**
Lot: **EBVNG-100**
Expiry: **2027-07-31**
IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0580-100	2026-01	2027-07-31
CONJ	1	KG0580-100	2026-01	2027-07-31
CONTROL -	1	NG0580-100	2026-01	2027-07-31
CUT OFF	1	CG0580-100	2026-01	2027-07-31
CONTROL +	1	PG0580-100	2026-01	2027-07-31
WASH BUF 20x	1	W-1132	2026-01	2028-01-31
SOLN STOP	1	S-1066	2025-12	2027-12-31
SUB TMB	1	T-079	2025-12	2028-12-10
DIL	1	D-498	2025-12	2027-12-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.000	< 0.200 < Cut-off
CUT OFF	0.558	0.150-1.300
CONTROL +	1.058	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	20.35	11.19 - 29.51

The target value and target range of CONTROL + may be used optionally.

The positive control is valid if the absorbance value of CONTROL + > CUT OFF.

Final Report issued: 2026-02-12

Approved by: Quality Control

Approved on: 2026-02-12

Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany

IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany