

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Epstein-Barr Virus (EBNA) IgG**

Prod.No.: **RE58741**

Lot: **EBVNG-101**

Expiry: **2027-12-31**

IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0580-101	2026-06	2027-12-31
CONJ	1	KG0580-101	2026-06	2027-12-31
CONTROL -	1	NG0580-101	2026-06	2027-12-31
CUT OFF	1	CG0580-101	2026-06	2027-12-31
CONTROL +	1	PG0580-101	2026-06	2027-12-31
WASH BUF 20x	1	W-1145	2026-06	2028-06-30
SOLN STOP	1	S-1076	2026-04	2028-04-30
SUB TMB	1	T-080	2026-02	2029-02-23
DIL	1	D-502	2026-04	2028-04-30


Component	OD Value	OD Validation Criteria
CONTROL -	0.013	< 0.200 < Cut-off
CUT OFF	0.595	0.150-1.300
CONTROL +	1.150	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	20.84	11.46 - 30.22

The target value and target range of **CONTROL +** may be used optionally.

The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2026-07-07

Approved by: 
Quality Control

Approved on: 2026-07-07



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany