

**GOLD  
STANDARD  
DIAGNOSTICS**



# Certificate of Analysis (CoA)

Product: **Epstein-Barr Virus (VCA) IgG**  
 Prod.No.: **RE58751**  
 Lot: **EBVG-216**  
 Expiry: **2026-10-31**  
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
<b>MTP</b>	1	MG0150-216	2025-04	2026-10-31
<b>CONJ</b>	1	KG0150-216	2025-04	2026-10-31
<b>CONTROL -</b>	1	NG0150-216	2025-04	2026-10-31
<b>CUT OFF</b>	1	CG0150-216	2025-04	2026-10-31
<b>CONTROL +</b>	1	PG0150-216	2025-04	2026-10-31
<b>WASH BUF 20x</b>	1	W-1109	2025-02	2027-02-28
<b>SOLN STOP</b>	1	S-1040	2025-02	2027-02-28
<b>SUB TMB</b>	1	T-077	2022-09	2027-03-31
<b>DIL</b>	1	D-490	2024-12	2026-12-31

Component	OD Value	OD Validation Criteria
<b>CONTROL -</b>	0.016	< 0.200 < Cut-off
<b>CUT OFF</b>	0.517	0.150-1.300
<b>CONTROL +</b>	1.058	> Cut-off


Component	Target Value [NTU]	Target Range [NTU]
<b>CONTROL +</b>	22.43	12.34 - 32.52


The target value and target range of **CONTROL +** may be used optionally.  
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2025-05-05

Approved by:   
 Quality Control

Approved on: 2025-05-06

 Gold Standard Diagnostics Frankfurt GmbH  
 Waldstrasse 23 A6  
 63128 Dietzenbach, Germany

 IBL International GmbH  
 Flughafenstrasse 52a, 22335 Hamburg  
 Germany